

High Technology, Inc. 109 Production Road, Walpole MA 02081	Document #	AV-DD-92123D-1507192
	Revision Level	B
	Prepared By	S. Titov
	Approved By	S. Titov
	Effective Date	10/1/2015
HTI Expected Ranges Chart DD-92123D		

Mission Control

Expected Range Chart

EXP: 05/2018
LOT: 1507192

ISE Analyzer		K			Na			Cl			Ca			pH			Li		
		Mean	Min	Max	Mean	Min	Max	Mean	Min	Max	Mean	Min	Max	Mean	Min	Max	Mean	Min	Max
LEVEL 1	E-Lyte 5	1.84	1.69	1.99	118	112	124	77	71	83	1.92	1.65	2.19	7.14	7.08	7.20			
	E-Lyte 5/Plus (NEW)	1.58	1.45	1.71	114	107	120	75	69	81	2.08	1.81	2.36	6.96	6.90	7.02			
LEVEL 2	E-Lyte 5	4.15	3.83	4.48	143	136	149	95	89	101	1.05	0.89	1.21	7.32	7.26	7.31			
	E-Lyte 5/Plus (NEW)	4.18	3.86	4.51	141	134	147	93	87	99	1.16	1.00	1.32	7.25	7.19	7.31			
LEVEL 3	E-Lyte 5	6.76	6.21	7.31	161	153	169	113	107	119	0.60	0.52	0.69	7.71	7.65	7.77			
	E-Lyte 5/Plus (NEW)	6.30	5.74	6.86	148	140	156	111	105	117	0.62	0.49	0.75	7.65	7.59	7.71			