

 High Technology, Inc. 109 Production Road, Walpole MA 02081	Document #	AV-DD-92123D-1512162
	Revision Level	B
	Prepared By	S. Titov
	Approved By	S. Titov
	Effective Date	10/1/2015
HTI Expected Ranges Chart DD-92123D		

Mission Control

Expected Range Chart

EXP: 7/30/2018
LOT: 1512162

ISE Analyzer		K			Na			Cl			Ca			pH			Li		
		Mean	Min	Max	Mean	Min	Max	Mean	Min	Max	Mean	Min	Max	Mean	Min	Max	Mean	Min	Max
LEVEL 1	E-Lyte 5	1.90	1.74	2.06	119	113	126	83	77	90	2.06	1.79	2.33	7.10	7.04	7.16			
	E-Lyte 5/Plus (NEW)	1.90	1.74	2.06	119	113	126	83	77	90	2.06	1.79	2.33	7.10	7.05	7.16			
LEVEL 2	E-Lyte 5	4.26	3.93	4.58	139	132	145	103	95	112	1.04	0.87	1.21	7.34	7.28	7.40			
	E-Lyte 5/Plus (NEW)	4.26	3.93	4.58	139	132	145	103	95	112	1.04	0.87	1.21	7.34	7.28	7.40			
LEVEL 3	E-Lyte 5	7.16	6.62	7.71	159	150	167	122	113	132	0.56	0.46	0.66	7.73	7.67	7.79			
	E-Lyte 5/Plus (NEW)	7.16	6.62	7.71	159	150	167	122	113	132	0.56	0.46	0.66	7.73	7.67	7.79			