

Hematology Calibrator₁ | System Specific Values₂

Open-vial stability 5 days₃



2019-01-16

350415-23
2018-04

LOT 83370194

Instrument₄
Abbott CELL-DYN® 3200

Parameter ₅	\bar{x}	+/-
WBC (WOC) 10 ⁹ /L	7.1	0.3
WBC (NOC) 10 ⁹ /L	7.0	0.3
RBC 10 ¹² /L	4.14	0.10
HGB g/dL	12.5	0.2
[HGB] g/L	125	2
MCV fL	81.3	2.0
PLT 10 ⁹ /L	242	15

Instrument₄
Abbott CELL-DYN® 3500 / 3700

Parameter ₅	\bar{x}	+/-
WBC (WOC) 10 ⁹ /L	7.0	0.3
WBC (WIC) 10 ⁹ /L	7.0	0.3
RBC 10 ¹² /L	4.15	0.10
HGB g/dL	12.6	0.2
[HGB] g/L	126	2
MCV fL	92.7	2.0
PLT 10 ⁹ /L	229	15

Instrument₄
Abbott CELL-DYN® Ruby®

Parameter ₅	\bar{x}	+/-
WBC (WOC) 10 ⁹ /L	7.1	0.3
WBC (NOC) 10 ⁹ /L	7.1	0.3
RBC 10 ¹² /L	4.19	0.10
HGB g/dL	12.5	0.2
[HGB] g/L	125	2
MCV fL	80.6	2.0
PLT 10 ⁹ /L	244	15

1 Hematology Calibrator

Etalon d'hématologie / Hämatologiekalibrator / Calibratore ematologico / Calibrador hematológico / Hematologikalibrator

2 System Specific Values

Valeurs spécifiques du système / Systemspezifische Werte / Valori specifici del sistema / Valores específicos del sistema / Systemspezifika värden

3 Open-vial stability 5 days

Stabilité en flacon ouvert 5 jours / Stabilität geöffneter Flaschen 5 tage / Stabilità della fiala aperta 5 giorni / Estabilidad de la cápsula abierta 5 días / Hållbarhet för öppen flaska 5 dagar

4 Instrument

Instrument / Gerät / Strumento / Instrumento / Instrument

5 Parameter

Paramètre / Parameter / Parametro / Parámetro / Parameter

\bar{x} Mean

Moyenne / Mittelwert / Media / Media / Medelvärde

+/- Expected Range

Intervalle escompté / Erwartungsbereich / Range previsto / Intervalo previsto / Förväntat intervall

[SI Units

Unités SI / SI- Einheiten / Unità SI / Unidades SI / SI-enheter

The brand and product names of the instruments are trademarks of their respective holders.

Les noms de marques et de produits des instruments appartiennent à leurs détenteurs respectifs. / Die Marken- und Produktnamen der Geräte sind Warenzeichen der jeweiligen Inhaber. / Le marche e i nomi degli strumenti sono marchi registrati dei rispettivi titolari. / Los nombres de marcas y productos de los instrumentos son marcas comerciales de sus titulares respectivos. / Instrumentmärken och produktnamn är varumärkta av respektive innehavare.



Streck
7002 S. 109 Street, La Vista, NE 68128 USA



MediMark® Europe
11, rue Emile Zola, BP 2332
38033 Grenoble Cedex 2, France